

Tender Number: 60135251
Date Created: Jun 25, 2008
Contact Person: Theresa Broussar
Telephone: 902-424-5524
Document Reference: 60135251

Tender Source:

TENDERS WEBSITE
WWW.GOV.NS.CA/TENDERS
6176 YOUNG ST
HALIFAX NS B3K 2A6

Bidders are responsible for ensuring that they are aware of and have complied with any Addenda by visiting the Procurement Website or contacting the Public Tenders Office.

Send Quotation To:

Public Tenders Office
6176 Young Street, Suite 200
Halifax, NS B3K 2A6
Ph. 902-424-3333, Fax 424-0622

Important Dates:

Closing Date: Jul 14, 2008
Closing Time: 10:00 AM

Bids are opened one half hour after tender closing

Deliver Goods\Services To:

DEPT. OF AGRICULTURE & FISHERIES
Truro NS B2N 5E3

For information on other tenders refer to:
www.gov.ns.ca/tenders

Delivery Requested By: 24/07/2008
(DD/MM/YYYY)

#10023892

CONTACT PERSON:
Kala Muthuswamy
902-896-2458

PLEASE STATE YOUR EARLIEST POSSIBLE DELIVERY.

You MUST submit current product brochure with this tender submission in order for your bid to be considered.

PRICES TO BE QUOTED TAX OUT ONLY

PLEASE COMPLETE THE UNIT, DELIVERY DATE, UNIT PRICE AND EXTENDED PRICE FIELDS

Item	Qty Unit	Material Description	Delivery date	ALL PRICES MUST BE EXTENDED AND TOTALLED	
				Unit Price	Extended Price

PRICES TO BE QUOTED TAX OUT ONLY

PLEASE COMPLETE THE UNIT, DELIVERY DATE, UNIT PRICE AND EXTENDED PRICE FIELDS

Item	Qty Unit	Material Description	Delivery date	ALL PRICES MUST BE EXTENDED AND TOTALLED	
				Unit Price	Extended Price
00001	1 EACH	Biological Safety Cabinet (four feet) 1 each Class II Type A2 Biological Safety Cabinet (four feet). Equivalent to Labconco Logic Class II. Specifications as follows: -must have a filter monitoring system consisting of an electronically commutated motor (ECM) that delivers a precise volume of air as required and automatically adjusts as filters load without relying on airflow sensors. -must have a system that allows the user to program start up and shut down operations. -must have a system that idles the blower when the sash is fully closed -must have built-in interval or elapsed timer for experiment monitoring fluorescent light. -must have touch pad control for manual activation of blower, light, timer, audible alarm mute and menu selection. -must have removable seamless stainless steel interior and work surface -must come with supply and exhaust 99.99% efficient HEPA filters. -must include 254nm UV lamp -must include two electrical duplex receptacles, located one on each side, with ground fault interruption and stainless steel splash covers -must come with a fully-closing, clear 1/4" tempered safety glass sash with two sash handles, counterbalanced, anti-racking mechanism, and 10 degree slope. -must have a sash opening viewing height of at least 50 cm -must have interior-mounted, line-of-sight LCD information center with "Filter Life Remaining" bar graph, status line for alarm conditions and alerts to warn when filter life diminishes to 20%, 10% and 0% -must come with an adjustable support base stand (30" or 36") -Equivalent to Labconco Logic Class II			

THE FOLLOWING INFORMATION MUST BE COMPLETED TO ENSURE TENDER ACCEPTANCE

TOTAL:

BIDDING COMPANY:

REPRESENTATIVE OF BIDDING COMPANY:

PRINT NAME:

PHONE#:

FAX#:

E-MAIL ADDRESS:

PO BOX:

CITY:

POSTAL CODE:

STREET:

CITY:

POSTAL CODE:

DELIVERY PROMISED:

TERMS:

FOB:

SIGNATURE:

All documents listed here can be found on our web site at www.gov.ns.ca/tenders/policy

If you do not have access to these documents via the Internet, please request a copy from Nova Scotia Procurement, 6176 Young St., PO Box 787, Halifax, NS B3J 2V2, phone (902)424-3333, fax (902)424-0622. Please request the document(s) by name.

NEED HELP? Suggestions and hints to help you reply to this tender are available from the above web page by selecting "Tendering Guides" then "Request for Quotations Completion Guide".

*******INSTRUCTIONS REGARDING THIS PURCHASE*******

By submitting a response to this tender, you acknowledge that you have read and complied with the applicable Nova Scotia Procurement documents. The following documents apply to this tender, and are available from the above web page by selecting "Terms & Conditions" then selecting the following:

Atlantic Standard Terms and Conditions (revised April 1, 2007)

#10023892